
Midmark M11 - v1.0.0
Total Cycles: 3427

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2023 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.5 F 0.0
2:00 178.1 F 0.2
4:00 181.0 F 3.3
6:00 203.0 F 7.9
8:00 220.1 F 13.0
10:00 227.9 F 7.9
12:00 240.5 F 13.0
14:00 239.4 F 10.6
16:00 236.6 F 8.7
18:00 251.1 F 15.7
20:00 262.8 F 22.4
22:00 272.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.6 F 28.8
2:00 271.6 F 28.0
3:00 272.5 F 28.9
4:00 272.2 F 28.5

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 227.6 F 0.0
10:00 229.0 F 0.0
15:00 230.0 F 0.0
20:00 231.3 F 0.0
25:00 232.0 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:32
CYCLE COMPLETE

06/22/2023 09:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____