
Midmark M11 - v1.0.5
Total Cycles: 3264

Name: RITTER2
Sterilizer ID: V2313228
12/28/2023 05:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.9 F 0.0
2:00 155.5 F 2.4
4:00 178.1 F 6.0
6:00 196.4 F 10.6
8:00 223.0 F 5.4
10:00 228.9 F 10.3
12:00 232.9 F 8.0
14:00 243.1 F 12.8
16:00 242.1 F 11.8
18:00 255.2 F 18.5
20:00 266.7 F 25.5
22:00 273.1 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.7 F 29.6
2:00 272.7 F 30.0
3:00 271.4 F 29.3
4:00 272.7 F 29.9

Min 271.3 F 29.3
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.8 F 0.0
5:00 251.6 F 0.0
10:00 197.1 F 0.0
15:00 190.1 F 0.0
20:00 188.6 F 0.0
25:00 186.0 F 0.0
30:00 183.1 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

12/28/2023 06:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____