
Midmark M11 - v1.0.5
Total Cycles: 167

Name: RITTER2
Sterilizer ID: V2313228
07/06/2021 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.1 F 0.0
2:00 172.1 F 3.1
4:00 186.0 F 7.5
6:00 205.5 F 12.7
8:00 222.8 F 7.6
10:00 234.9 F 12.8
12:00 237.1 F 10.6
14:00 236.0 F 9.2
16:00 249.7 F 15.8
18:00 261.1 F 22.3
20:00 271.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.6 F 29.8
2:00 271.6 F 29.1
3:00 272.3 F 30.0
4:00 272.2 F 29.5

Min 271.3 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 226.8 F 0.0
10:00 232.0 F 0.0
15:00 226.3 F 0.0
20:00 226.3 F 0.0
25:00 227.8 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

07/06/2021 11:39 AM
ACCEPTED: X
REJECTED:
Operator: U02