
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001124
10:34 20/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:11
PRESSURIZING 4:13
STERILIZING 5:32
136.0 C 220kPa 5:32
136.0 C 221kPa 6:02
136.2 C 223kPa 6:32
136.4 C 226kPa 7:02
136.4 C 226kPa 7:32
136.4 C 226kPa 8:02
136.4 C 226kPa 8:32
136.4 C 226kPa 9:02
136.4 C 226kPa 9:02
VENTING 9:02
AIR DRYING 10:03
CYCLE COMPLETE 17:15

Digital Signature #
66108CDB02F2355D
