
Midmark M11 - v1.0.5
Total Cycles: 1529

Name: RITTER2
Sterilizer ID: V2313228
09/22/2022 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.1 F 0.0
2:00 189.1 F 4.0
4:00 197.0 F 9.2
6:00 221.5 F 7.8
8:00 230.3 F 10.6
10:00 232.5 F 8.5
12:00 239.3 F 10.2
14:00 247.7 F 14.9
16:00 260.6 F 22.0
18:00 271.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.1 F 30.1
2:00 271.7 F 29.4
3:00 271.6 F 29.9
4:00 272.3 F 29.8

Min 271.1 F 29.2
Max 272.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 239.4 F 0.0
10:00 228.0 F 0.0
15:00 231.5 F 0.0
20:00 227.2 F 0.0
25:00 229.1 F 0.0
Dry Cycle Aborted

FILLING: 2:24
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:15
DRYING: 27:19
TOTAL CYCLE: 00:55:22

09/22/2022 05:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____