
Midmark M11 - v1.0.5
Total Cycles: 3140

Name: RITTER2
Sterilizer ID: V2313228
11/17/2023 03:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.2 F 0.0
2:00 98.3 F 0.2
4:00 119.7 F 1.8
6:00 144.6 F 4.6
8:00 168.4 F 8.0
10:00 188.2 F 11.8
12:00 211.9 F 4.8
14:00 215.7 F 8.7
16:00 228.2 F 13.2
18:00 235.4 F 9.8
20:00 237.3 F 9.7
22:00 246.2 F 13.9
24:00 257.4 F 20.1
26:00 267.9 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.7
1:00 272.7 F 29.8
2:00 272.8 F 29.9
3:00 272.8 F 30.1
4:00 272.3 F 30.2

Min 271.6 F 29.6
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 202.2 F 0.0
10:00 189.3 F 0.0
15:00 186.7 F 0.0
20:00 186.0 F 0.0
25:00 184.7 F 0.0
30:00 183.3 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 27:54
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:06:48
CYCLE COMPLETE

11/17/2023 04:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____