
Midmark M11 - v1.0.5
Total Cycles: 2089

Name: RITTER2
Sterilizer ID: V2313228
02/23/2023 05:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.3 F 0.0
2:00 185.0 F 2.9
4:00 203.3 F 7.9
6:00 223.8 F 11.8
8:00 231.6 F 10.0
10:00 233.7 F 8.2
12:00 240.9 F 11.1
14:00 246.9 F 14.1
16:00 260.1 F 21.5
18:00 271.9 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.4 F 29.4
2:00 271.7 F 29.4
3:00 272.4 F 29.5
4:00 271.3 F 29.0

Min 271.3 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 212.5 F 0.0
10:00 227.6 F 0.0
15:00 231.5 F 0.0
20:00 224.6 F 0.0
25:00 221.7 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:33
CYCLE COMPLETE

02/23/2023 06:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____