

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 001601  
13:30 11/19/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:15  
PRESSURIZING 6:59  
STERILIZING 8:37  
136.1 C 220kPa 8:37  
136.2 C 221kPa 9:07  
135.9 C 218kPa 9:37  
136.4 C 222kPa 10:07  
136.3 C 221kPa 10:38  
136.0 C 219kPa 11:08  
136.5 C 222kPa 11:38  
136.4 C 220kPa 12:07  
VENTING 12:07  
AIR DRYING 13:13  
CYCLE COMPLETE 43:13  
-----

Digital Signature #  
D482F428D185CA97  
-----