

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000639  
9:36 09/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:11  
PRESSURIZING 3:16  
STERILIZING 4:54  
136.4 C 230kPa 4:54  
136.3 C 232kPa 5:24  
136.4 C 230kPa 5:54  
136.2 C 230kPa 6:24  
136.2 C 230kPa 6:54  
136.2 C 230kPa 7:24  
136.2 C 231kPa 7:54  
136.2 C 231kPa 8:24  
136.2 C 231kPa 8:24  
VENTING 8:24  
AIR DRYING 9:17  
CYCLE COMPLETE 24:17  
-----

Digital Signature #  
F39226A9E115213E  
-----