
Midmark M11 - v1.0.0
Total Cycles: 1513

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2022 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.3 F 0.0
2:00 118.9 F 0.8
4:00 141.1 F 3.3
6:00 166.9 F 6.8
8:00 190.5 F 11.0
10:00 207.6 F 3.7
12:00 220.8 F 8.6
14:00 232.6 F 12.9
16:00 233.3 F 9.1
18:00 240.1 F 10.6
20:00 243.8 F 12.6
22:00 255.7 F 18.6
24:00 266.1 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 272.3 F 29.3
2:00 272.1 F 28.8
3:00 271.8 F 29.0
4:00 272.7 F 29.3

Min 271.5 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 207.5 F 0.0
10:00 224.7 F 0.0
15:00 229.3 F 0.0
20:00 229.2 F 0.0
25:00 228.6 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 25:34
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:03:41
CYCLE COMPLETE

01/12/2022 11:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____