

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1029

Name: RITTER 1  
Sterilizer ID: V2178198  
07/19/2021 12:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.3 F 0.0  
2:00 155.4 F 1.4  
4:00 161.0 F 3.8  
6:00 183.4 F 7.2  
8:00 201.9 F 11.3  
10:00 222.6 F 5.7  
12:00 233.5 F 10.9  
14:00 228.9 F 7.2  
16:00 243.6 F 13.2  
18:00 241.9 F 11.5  
20:00 254.0 F 17.6  
22:00 264.6 F 24.0  
24:00 270.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 271.6 F 29.0  
2:00 271.7 F 28.9  
3:00 272.2 F 29.1  
4:00 272.5 F 29.2

Min 271.5 F 28.9  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
5:00 236.8 F 0.0  
10:00 219.2 F 0.0  
15:00 203.1 F 0.0  
20:00 187.8 F 0.0  
25:00 191.6 F 0.0  
30:00 211.3 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 24:36  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:03:17  
CYCLE COMPLETE

07/19/2021 01:38 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01