

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3359

Name: RITTER2  
Sterilizer ID: V2313228  
02/06/2024 11:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.8 F 0.0  
2:00 127.8 F 0.0  
4:00 131.3 F 1.5  
6:00 154.7 F 4.7  
8:00 180.9 F 8.9  
10:00 202.2 F 13.5  
12:00 219.1 F 6.7  
14:00 231.9 F 11.7  
16:00 232.8 F 8.7  
18:00 240.5 F 10.7  
20:00 245.7 F 13.5  
22:00 258.0 F 20.1  
24:00 269.0 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 272.5 F 29.6  
2:00 271.9 F 29.0  
3:00 272.9 F 29.5  
4:00 271.9 F 29.4

Min 271.0 F 28.5  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
5:00 218.6 F 0.0  
10:00 196.3 F 0.0  
15:00 195.1 F 0.0  
20:00 192.5 F 0.0  
25:00 190.2 F 0.0  
30:00 188.0 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 25:11  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:03:19  
CYCLE COMPLETE

02/06/2024 12:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_