
Midmark M11 - v1.0.0
Total Cycles: 1574

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2022 02:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.6 F 0.0
2:00 165.2 F 3.2
4:00 173.9 F 7.1
6:00 204.2 F 12.4
8:00 218.2 F 6.2
10:00 232.4 F 11.2
12:00 232.2 F 7.8
14:00 244.0 F 12.4
16:00 244.7 F 12.4
18:00 257.3 F 19.1
20:00 268.6 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.2 F 28.8
2:00 272.3 F 28.6
3:00 271.4 F 28.2
4:00 272.7 F 29.1

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 232.1 F 0.0
10:00 224.3 F 0.0
15:00 223.1 F 0.0
20:00 222.7 F 0.0
25:00 224.7 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

01/31/2022 03:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____