
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002362
9:03 05/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:13
STERILIZING 4:36
136.4 C 220kPa 4:36
136.4 C 222kPa 5:06
136.4 C 220kPa 5:36
136.4 C 221kPa 6:06
136.4 C 221kPa 6:36
136.4 C 223kPa 7:06
136.1 C 220kPa 7:36
136.4 C 223kPa 8:06
136.4 C 223kPa 8:06
VENTING 8:06
AIR DRYING 9:00
CYCLE COMPLETE 9:01

Digital Signature #
2B4ABF5910F03C37
