

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5147

Name: RITTER 1  
Sterilizer ID: V2178198  
12/02/2024 11:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 106.8 F 0.0  
2:00 107.7 F 0.0  
4:00 119.1 F 0.1  
6:00 139.3 F 0.2  
8:00 166.7 F 0.5  
10:00 187.5 F 1.0  
12:00 205.0 F 4.3  
14:00 222.9 F 8.1  
16:00 232.8 F 12.0  
18:00 231.6 F 8.0  
20:00 242.9 F 13.2  
22:00 241.6 F 11.2  
24:00 238.3 F 9.5  
26:00 251.3 F 15.6  
28:00 261.4 F 21.4  
30:00 270.5 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.0 F 28.5  
2:00 272.7 F 28.9  
3:00 272.5 F 28.6  
4:00 272.1 F 28.3

Min 271.5 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 200.9 F 0.0  
10:00 208.6 F 0.0  
15:00 210.4 F 0.0  
20:00 218.5 F 0.0  
25:00 225.4 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:01  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:09:32  
CYCLE COMPLETE

12/02/2024 12:59 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_