
Midmark M11 - v1.0.0
Total Cycles: 3707

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2023 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.1 F 0.0
2:00 145.2 F 2.4
4:00 166.7 F 5.5
6:00 185.9 F 8.9
8:00 201.0 F 13.0
10:00 220.1 F 5.8
12:00 224.6 F 10.1
14:00 233.5 F 8.7
16:00 240.1 F 11.3
18:00 237.0 F 9.2
20:00 249.8 F 15.1
22:00 260.6 F 21.1
24:00 270.4 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 271.9 F 28.7
2:00 272.0 F 28.8
3:00 272.0 F 29.1
4:00 272.1 F 29.3

Min 271.5 F 28.6
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 240.2 F 0.0
10:00 224.9 F 0.0
15:00 230.2 F 0.0
20:00 219.7 F 0.0
25:00 216.0 F 0.0
30:00 239.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:29
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:04:06
CYCLE COMPLETE

08/30/2023 03:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____