
Midmark M11 - v1.0.0
Total Cycles: 2440

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2022 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 218.9 F 0.0
2:00 214.8 F 4.8
4:00 217.8 F 10.1
6:00 228.2 F 6.0
8:00 233.9 F 11.1
10:00 234.8 F 8.5
12:00 239.7 F 10.3
14:00 246.8 F 13.8
16:00 259.2 F 20.5
18:00 270.2 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.2
2:00 271.9 F 28.6
3:00 271.6 F 28.7
4:00 272.5 F 29.2

Min 271.3 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.4 F 0.0
5:00 233.6 F 0.0
10:00 228.7 F 0.0
15:00 228.7 F 0.0
20:00 227.1 F 0.0
25:00 228.0 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 2:27
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

10/17/2022 11:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____