
Midmark M11 - v1.0.0
Total Cycles: 3202

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2023 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.6 F 0.0
2:00 148.6 F 0.0
4:00 176.5 F 1.8
6:00 196.4 F 6.0
8:00 210.5 F 10.7
10:00 220.7 F 5.4
12:00 233.8 F 10.6
14:00 232.7 F 7.5
16:00 245.0 F 12.8
18:00 244.3 F 12.4
20:00 257.1 F 19.1
22:00 268.4 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.9
2:00 271.5 F 28.3
3:00 272.6 F 29.0
4:00 271.8 F 28.3

Min 271.4 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 218.8 F 0.0
10:00 230.7 F 0.0
15:00 223.3 F 0.0
20:00 227.3 F 0.0
25:00 228.0 F 0.0
30:00 214.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:12
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

05/04/2023 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____