
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008341
12:45 24/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:09
STERILIZING 4:14
136.0 C 220kPa 4:14
136.1 C 219kPa 4:45
136.2 C 221kPa 5:15
136.4 C 224kPa 5:45
136.4 C 225kPa 6:15
136.2 C 221kPa 6:45
136.4 C 223kPa 7:15
136.4 C 221kPa 7:45
136.4 C 221kPa 7:45
VENTING 7:45
AIR DRYING 8:41
CYCLE COMPLETE 23:35

Digital Signature #
B420CBBA45D55768
