
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004240
14:42 25/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:16
STERILIZING 3:09
136.0 C 220kPa 3:09
136.3 C 225kPa 3:39
136.4 C 225kPa 4:09
136.4 C 225kPa 4:39
136.4 C 225kPa 5:09
136.3 C 224kPa 5:39
136.2 C 219kPa 6:09
136.4 C 225kPa 6:39
136.4 C 225kPa 6:39
VENTING 6:40
AIR DRYING 8:45
CYCLE COMPLETE 8:53

Digital Signature #
1CAFF3C5F14B11F6
