
Midmark M11 - v1.0.0
Total Cycles: 2139

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2022 02:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 187.9 F 3.9
4:00 193.1 F 8.4
6:00 215.5 F 11.5
8:00 225.6 F 8.5
10:00 237.8 F 11.8
12:00 239.1 F 10.9
14:00 236.7 F 8.9
16:00 250.5 F 15.5
18:00 261.6 F 21.9
20:00 272.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.0
2:00 271.5 F 28.2
3:00 272.7 F 29.2
4:00 272.0 F 28.6

Min 271.2 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 237.3 F 0.0
10:00 232.4 F 0.0
15:00 227.1 F 0.0
20:00 222.2 F 0.0
25:00 224.4 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:41
CYCLE COMPLETE

07/15/2022 03:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____