

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002415  
9:17 13/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:27  
PRESSURIZING 3:29  
STERILIZING 4:37  
136.4 C 224kPa 4:37  
136.4 C 224kPa 5:07  
136.4 C 224kPa 5:37  
136.4 C 219kPa 6:07  
135.9 C 220kPa 6:37  
136.4 C 224kPa 7:07  
136.4 C 223kPa 7:37  
136.0 C 222kPa 8:07  
136.0 C 222kPa 8:07  
VENTING 8:07  
AIR DRYING 8:57  
CYCLE COMPLETE 14:38  
-----

Digital Signature #  
B3F7CF8E48962375  
-----