
Midmark M11 - v1.0.0
Total Cycles: 2294

Name: RITTER 1
Sterilizer ID: V2178198
09/07/2022 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.2 F 0.0
2:00 75.3 F 0.0
4:00 86.4 F 0.8
6:00 120.9 F 3.1
8:00 150.1 F 6.0
10:00 173.7 F 9.2
12:00 194.0 F 12.8
14:00 207.3 F 3.6
16:00 218.3 F 6.9
18:00 228.0 F 10.2
20:00 233.5 F 9.7
22:00 236.4 F 10.0
24:00 237.1 F 8.8
26:00 245.7 F 13.1
28:00 256.7 F 19.0
30:00 266.7 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 271.8 F 28.9
2:00 272.4 F 28.9
3:00 271.4 F 28.4
4:00 272.5 F 29.2

Min 271.3 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 207.1 F 0.0
10:00 216.1 F 0.0
15:00 219.1 F 0.0
20:00 229.6 F 0.0
25:00 237.3 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:29
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:28
CYCLE COMPLETE

09/07/2022 09:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____