
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003984
10:13 11/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:01
STERILIZING 3:04
136.4 C 220kPa 3:04
136.4 C 225kPa 3:34
136.4 C 224kPa 4:04
136.1 C 222kPa 4:34
136.4 C 225kPa 5:04
136.4 C 224kPa 5:34
136.4 C 224kPa 6:04
136.4 C 223kPa 6:34
136.4 C 223kPa 6:34
VENTING 6:35
AIR DRYING 8:29
CYCLE COMPLETE 10:40

Digital Signature #
98AE2611AF81C553
