

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4439

Name: RITTER 1  
Sterilizer ID: V2178198  
04/09/2024 12:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.1 F 0.0  
2:00 111.5 F 0.9  
4:00 118.6 F 3.0  
6:00 146.3 F 5.7  
8:00 181.4 F 9.1  
10:00 195.5 F 13.1  
12:00 211.3 F 4.5  
14:00 218.6 F 8.0  
16:00 229.0 F 11.8  
18:00 230.1 F 7.5  
20:00 241.2 F 12.4  
22:00 239.2 F 10.3  
24:00 250.9 F 15.9  
26:00 261.3 F 21.8  
28:00 271.2 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.4  
1:00 272.4 F 29.2  
2:00 272.6 F 29.1  
3:00 271.7 F 29.0  
4:00 272.5 F 29.1  
  
Min 271.0 F 28.2  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 220.5 F 0.0  
10:00 217.5 F 0.0  
15:00 212.9 F 0.0  
20:00 215.3 F 0.0  
25:00 228.2 F 0.0  
30:00 236.1 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 28:47  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:07:33  
CYCLE COMPLETE

04/09/2024 01:10 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_