
Midmark M11 - v1.0.0
Total Cycles: 5332

Name: RITTER 1
Sterilizer ID: V2178198
02/11/2025 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 70.0 F 0.3
4:00 80.5 F 1.7
6:00 118.2 F 3.8
8:00 145.4 F 6.2
10:00 165.9 F 9.0
12:00 186.1 F 12.2
14:00 200.8 F 2.2
16:00 210.1 F 5.9
18:00 222.0 F 9.4
20:00 232.7 F 13.3
22:00 233.2 F 8.5
24:00 243.0 F 13.3
26:00 241.7 F 11.0
28:00 252.8 F 16.7
30:00 262.7 F 22.5
32:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.4 F 29.1
2:00 271.7 F 28.7
3:00 272.3 F 29.1
4:00 271.6 F 28.7

Min 271.6 F 28.6
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 195.6 F 0.0
10:00 192.5 F 0.0
15:00 182.7 F 0.0
20:00 180.5 F 0.0
25:00 184.8 F 0.0
30:00 194.9 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 33:24
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:11:51
CYCLE COMPLETE

02/11/2025 11:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____