

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2273

Name: RITTER2  
Sterilizer ID: V2313228  
04/20/2023 08:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
2:00 177.0 F 2.4  
4:00 176.0 F 6.0  
6:00 203.0 F 11.1  
8:00 222.2 F 5.9  
10:00 232.0 F 11.4  
12:00 234.8 F 8.8  
14:00 239.8 F 10.1  
16:00 247.3 F 14.1  
18:00 259.5 F 20.8  
20:00 270.5 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.8 F 29.5  
2:00 271.4 F 28.7  
3:00 272.7 F 29.5  
4:00 272.3 F 29.5  
  
Min 270.9 F 28.6  
Max 273.1 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.1 F 0.0  
5:00 226.1 F 0.0  
10:00 235.6 F 0.0  
15:00 231.2 F 0.0  
20:00 226.3 F 0.0  
25:00 227.1 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 20:51  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:34  
CYCLE COMPLETE

04/20/2023 09:28 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_