
Midmark M11 - v1.0.0
Total Cycles: 4402

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2024 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.5 F 0.0
2:00 104.1 F 0.8
4:00 117.2 F 2.8
6:00 146.0 F 5.6
8:00 171.4 F 9.1
10:00 192.6 F 12.9
12:00 208.8 F 3.9
14:00 221.0 F 7.4
16:00 230.8 F 11.3
18:00 229.8 F 6.6
20:00 242.1 F 12.0
22:00 239.0 F 9.6
24:00 251.1 F 15.5
26:00 261.6 F 21.4
28:00 271.2 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.4 F 28.6
2:00 271.2 F 27.9
3:00 272.1 F 28.7
4:00 272.3 F 28.5

Min 271.0 F 27.9
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.2 F 0.0
5:00 236.7 F 0.0
10:00 221.9 F 0.0
15:00 212.2 F 0.0
20:00 216.1 F 0.0
25:00 218.8 F 0.0
30:00 204.1 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 28:44
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:07:18
CYCLE COMPLETE

03/25/2024 01:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____