
Midmark M11 - v1.0.5
Total Cycles: 2777

Name: RITTER2
Sterilizer ID: V2313228
08/16/2023 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.9 F 0.0
2:00 163.9 F 0.0
4:00 185.4 F 0.0
6:00 203.8 F 0.0
8:00 217.4 F 2.9
10:00 227.3 F 6.9
12:00 238.1 F 11.3
14:00 234.9 F 8.8
16:00 241.7 F 11.1
18:00 245.7 F 12.7
20:00 246.3 F 13.5
22:00 258.3 F 20.0
24:00 268.7 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.4 F 29.1
2:00 271.6 F 28.8
3:00 272.3 F 29.0
4:00 271.4 F 28.7

Min 271.4 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 234.3 F 0.0
10:00 219.9 F 0.0
15:00 218.8 F 0.0
20:00 226.3 F 0.0
25:00 224.3 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:12
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:35
CYCLE COMPLETE

08/16/2023 12:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____