

-----  
Midmark M11 - v1.0.5  
Total Cycles: 204

Name: RITTER2  
Sterilizer ID: V2313228  
07/14/2021 07:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.4 F 0.0  
2:00 175.8 F 1.6  
4:00 177.1 F 5.3  
6:00 192.6 F 9.9  
8:00 218.5 F 6.5  
10:00 226.8 F 10.0  
12:00 233.7 F 8.7  
14:00 241.9 F 13.1  
16:00 242.2 F 12.0  
18:00 254.5 F 18.4  
20:00 265.3 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.1 F 29.6  
2:00 271.9 F 29.9  
3:00 271.6 F 29.2  
4:00 272.2 F 29.9

Min 271.3 F 29.1  
Max 272.4 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.8 F 0.0  
5:00 214.1 F 0.0  
10:00 220.9 F 0.0  
15:00 223.3 F 0.0  
20:00 224.0 F 0.0  
25:00 196.5 F 0.0  
Dry Cycle Aborted

FILLING: 1:40  
HEATING: 21:48  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 27:20  
TOTAL CYCLE: 00:57:58

07/14/2021 08:36 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01