

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001407  
13:53 27/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:36  
STERILIZING 3:41  
136.4 C 223kPa 3:41  
136.4 C 226kPa 4:11  
136.4 C 225kPa 4:41  
136.2 C 224kPa 5:11  
136.4 C 225kPa 5:41  
136.3 C 223kPa 6:11  
136.4 C 224kPa 6:41  
136.4 C 222kPa 7:11  
136.4 C 222kPa 7:11  
VENTING 7:11  
AIR DRYING 8:01  
CYCLE COMPLETE 23:01  
-----

Digital Signature #  
21602A2A6C757CC4  
-----