

-----  
Midmark M11 - v1.0.5  
Total Cycles: 316

Name: RITTER2  
Sterilizer ID: V2313228  
08/27/2021 12:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	129.9 F	0.0
2:00	129.9 F	1.4
4:00	137.6 F	4.1
6:00	163.7 F	7.7
8:00	187.2 F	12.3
10:00	206.3 F	5.1
12:00	220.9 F	9.7
14:00	230.8 F	9.4
16:00	237.1 F	12.1
18:00	239.3 F	10.7
20:00	252.1 F	17.1
22:00	262.2 F	23.6
24:00	272.5 F	30.6

STERILIZING

mm:ss	Degrees	PSI
0:00	271.6 F	29.8
1:00	272.8 F	30.6
2:00	272.4 F	30.0
3:00	271.8 F	29.9
4:00	272.7 F	30.6

Min 271.6 F 29.6  
Max 273.1 F 30.6

VENTING CHAMBER

DRYING

Drying Start

mm:ss	Degrees	PSI
0:00	178.7 F	0.0
5:00	221.7 F	0.0
10:00	232.4 F	0.0
15:00	216.9 F	0.0
20:00	206.4 F	0.0
25:00	224.7 F	0.0
30:00	214.8 F	0.0

Drying Complete

FILLING: 1:06  
HEATING: 24:28  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:02:43  
CYCLE COMPLETE

08/27/2021 01:20 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_