

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2933

Name: RITTER 1  
Sterilizer ID: V2178198  
02/22/2023 04:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.3 F 0.0  
2:00 164.5 F 3.9  
4:00 185.6 F 8.2  
6:00 204.4 F 13.3  
8:00 221.3 F 7.6  
10:00 234.5 F 12.9  
12:00 235.8 F 9.8  
14:00 235.7 F 8.4  
16:00 248.9 F 14.7  
18:00 261.3 F 21.7  
20:00 272.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.4 F 29.0  
2:00 271.4 F 28.6  
3:00 272.4 F 29.0  
4:00 271.3 F 28.4  
  
Min 271.2 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 215.3 F 0.0  
10:00 233.8 F 0.0  
15:00 234.8 F 0.0  
20:00 225.7 F 0.0  
25:00 221.9 F 0.0  
30:00 219.0 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 20:25  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:59  
CYCLE COMPLETE

02/22/2023 05:35 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_