

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2016

Name: RITTER 1  
Sterilizer ID: V2178198  
06/09/2022 10:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.1 F 0.0  
2:00 184.1 F 3.7  
4:00 208.6 F 8.3  
6:00 225.4 F 9.0  
8:00 229.7 F 8.8  
10:00 236.7 F 10.3  
12:00 241.2 F 11.5  
14:00 238.1 F 9.7  
16:00 252.1 F 16.5  
18:00 264.0 F 23.3  
20:00 271.3 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.7 F 29.2  
2:00 271.8 F 28.4  
3:00 272.0 F 28.9  
4:00 272.6 F 29.0  
  
Min 271.2 F 28.2  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
5:00 200.7 F 0.0  
10:00 211.2 F 0.0  
15:00 213.0 F 0.0  
20:00 193.5 F 0.0  
25:00 187.3 F 0.0  
30:00 188.6 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 20:03  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:56  
CYCLE COMPLETE

06/09/2022 11:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_