
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000817
7:31 13/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:33
STERILIZING 3:48
136.5 C 223kPa 3:48
136.3 C 223kPa 4:18
136.1 C 220kPa 4:48
136.4 C 221kPa 5:18
136.4 C 220kPa 5:48
136.2 C 223kPa 6:18
136.0 C 222kPa 6:48
136.2 C 218kPa 7:18
136.2 C 218kPa 7:18
VENTING 7:18
AIR DRYING 8:10
CYCLE COMPLETE 8:17

Digital Signature #
C6F37CF2F0B76186
