
Midmark M11 - v1.0.0
Total Cycles: 2005

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2022 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.6 F 0.0
2:00 172.7 F 4.0
4:00 190.3 F 8.8
6:00 222.0 F 8.7
8:00 230.8 F 8.7
10:00 237.2 F 10.8
12:00 239.5 F 11.1
14:00 237.2 F 9.2
16:00 251.5 F 16.1
18:00 263.5 F 23.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.0 F 28.9
2:00 272.5 F 28.8
3:00 271.2 F 28.3
4:00 272.6 F 29.2

Min 271.1 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 220.6 F 0.0
10:00 237.3 F 0.0
15:00 244.2 F 0.0
20:00 229.5 F 0.0
25:00 222.1 F 0.0
30:00 218.0 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

06/07/2022 04:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____