
Midmark M11 - v1.0.5
Total Cycles: 3029

Name: RITTER2
Sterilizer ID: V2313228
10/19/2023 09:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 170.4 F 1.8
4:00 182.1 F 6.0
6:00 205.7 F 11.4
8:00 222.7 F 6.2
10:00 235.6 F 11.6
12:00 234.8 F 9.1
14:00 237.7 F 9.7
16:00 247.9 F 14.5
18:00 260.2 F 21.4
20:00 271.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.2 F 29.6
2:00 271.9 F 29.0
3:00 272.1 F 29.6
4:00 272.1 F 29.2

Min 271.2 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.7 F 0.0
5:00 235.8 F 0.0
10:00 194.9 F 0.0
15:00 192.0 F 0.0
20:00 189.9 F 0.0
25:00 189.5 F 0.0
30:00 191.2 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

10/19/2023 10:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____