

-----  
Midmark M11 - v1.0.5  
Total Cycles: 610

Name: RITTER2  
Sterilizer ID: V2313228  
12/08/2021 09:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.4 F 0.0  
2:00 106.5 F 0.8  
4:00 121.3 F 3.0  
6:00 146.9 F 6.5  
8:00 175.6 F 11.1  
10:00 204.5 F 3.2  
12:00 218.9 F 7.6  
14:00 231.3 F 12.3  
16:00 234.0 F 9.3  
18:00 238.7 F 9.9  
20:00 246.8 F 14.1  
22:00 258.9 F 20.8  
24:00 270.0 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.1  
1:00 272.4 F 29.4  
2:00 272.8 F 29.9  
3:00 271.7 F 29.0  
4:00 272.7 F 29.6

Min 271.7 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.1 F 0.0  
5:00 212.8 F 0.0  
10:00 205.8 F 0.0  
15:00 202.3 F 0.0  
20:00 200.4 F 0.0  
25:00 199.2 F 0.0  
30:00 195.6 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 24:54  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 01:02:43  
CYCLE COMPLETE

12/08/2021 10:37 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02