

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3675

Name: RITTER 1  
Sterilizer ID: V2178198  
08/18/2023 10:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.5 F 0.0  
2:00 123.8 F 1.0  
4:00 140.0 F 3.2  
6:00 161.9 F 5.9  
8:00 180.1 F 9.1  
10:00 197.0 F 12.8  
12:00 213.6 F 4.8  
14:00 222.5 F 8.8  
16:00 234.4 F 13.1  
18:00 236.2 F 9.4  
20:00 239.6 F 10.5  
22:00 244.9 F 12.7  
24:00 256.0 F 18.5  
26:00 266.2 F 24.7  
28:00 271.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 271.8 F 28.8  
2:00 272.0 F 28.9  
3:00 271.9 F 29.2  
4:00 271.9 F 29.4

Min 271.4 F 28.8  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 228.7 F 0.0  
10:00 209.4 F 0.0  
15:00 211.5 F 0.0  
20:00 211.4 F 0.0  
25:00 233.8 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 28:29  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:06:59  
CYCLE COMPLETE

08/18/2023 11:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_