

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2745

Name: RITTER 1  
Sterilizer ID: V2178198  
01/05/2023 04:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.9 F 0.0  
2:00 154.7 F 0.0  
4:00 181.3 F 1.8  
6:00 202.6 F 6.2  
8:00 218.6 F 10.9  
10:00 223.6 F 5.9  
12:00 237.3 F 11.3  
14:00 234.9 F 8.7  
16:00 239.8 F 10.1  
18:00 247.4 F 14.0  
20:00 259.7 F 20.7  
22:00 270.6 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.3  
1:00 272.4 F 29.1  
2:00 271.9 F 28.4  
3:00 272.2 F 29.0  
4:00 272.1 F 28.6

Min 271.3 F 28.1  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.1 F 0.0  
5:00 216.8 F 0.0  
10:00 236.0 F 0.0  
15:00 230.1 F 0.0  
20:00 229.8 F 0.0  
25:00 230.7 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 22:42  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:52  
CYCLE COMPLETE

01/05/2023 05:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_