
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001600
7:48 20/05/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:04
STERILIZING 4:03
136.4 C 221kPa 4:03
136.4 C 223kPa 4:33
136.1 C 219kPa 5:03
136.2 C 221kPa 5:33
136.4 C 222kPa 6:03
136.3 C 222kPa 6:33
136.3 C 221kPa 7:03
136.3 C 223kPa 7:33
136.3 C 223kPa 7:33
VENTING 7:33
AIR DRYING 8:24
CYCLE COMPLETE 8:28

Digital Signature #
5A44398DBCEBC8AB
