
Midmark M11 - v1.0.5
Total Cycles: 3083

Name: RITTER2
Sterilizer ID: V2313228
11/03/2023 02:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.5 F 0.0
2:00 104.5 F 0.3
4:00 126.0 F 2.3
6:00 158.5 F 6.0
8:00 185.5 F 10.2
10:00 210.4 F 5.2
12:00 220.9 F 8.7
14:00 232.3 F 12.6
16:00 235.6 F 10.4
18:00 235.0 F 8.3
20:00 248.7 F 15.1
22:00 260.6 F 21.9
24:00 271.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.9 F 29.9
2:00 271.8 F 29.2
3:00 271.6 F 29.6
4:00 272.9 F 29.8

Min 271.1 F 28.9
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 203.8 F 0.0
10:00 192.4 F 0.0
15:00 189.0 F 0.0
20:00 187.3 F 0.0
25:00 186.2 F 0.0
30:00 186.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:36
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:02:50
CYCLE COMPLETE

11/03/2023 03:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____