

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4061

Name: RITTER 1  
Sterilizer ID: V2178198  
11/27/2023 03:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.0 F 0.0  
2:00 174.9 F 2.9  
4:00 182.0 F 7.1  
6:00 204.0 F 12.3  
8:00 222.3 F 6.3  
10:00 232.4 F 11.1  
12:00 231.9 F 7.7  
14:00 245.0 F 12.7  
16:00 243.9 F 12.3  
18:00 256.1 F 18.7  
20:00 267.2 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.2 F 28.6  
2:00 272.3 F 29.0  
3:00 271.2 F 28.3  
4:00 272.4 F 28.9  
  
Min 271.1 F 28.2  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.3 F 0.0  
5:00 229.7 F 0.0  
10:00 211.2 F 0.0  
15:00 209.9 F 0.0  
20:00 211.5 F 0.0  
25:00 208.9 F 0.0  
30:00 207.8 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 21:24  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:00:20  
CYCLE COMPLETE

11/27/2023 04:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_