

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001401  
9:28 09/08/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:54  
PRESSURIZING 5:23  
STERILIZING 6:46  
136.0 C 220kPa 6:46  
136.2 C 221kPa 7:16  
136.1 C 221kPa 7:46  
136.4 C 221kPa 8:16  
136.4 C 222kPa 8:46  
136.2 C 222kPa 9:16  
136.1 C 222kPa 9:46  
136.2 C 221kPa 10:16  
VENTING 10:16  
AIR DRYING 11:15  
CYCLE COMPLETE 11:21  
-----

Digital Signature #  
624CB1B8D550EFB7  
-----