
Midmark M11 - v1.0.5
Total Cycles: 648

Name: RITTER2
Sterilizer ID: V2313228
12/16/2021 04:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.6 F 0.0
2:00 168.1 F 3.3
4:00 181.8 F 7.6
6:00 200.2 F 12.5
8:00 218.2 F 6.3
10:00 230.7 F 11.3
12:00 233.8 F 8.8
14:00 240.5 F 10.6
16:00 247.4 F 14.5
18:00 259.7 F 21.3
20:00 270.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.5 F 29.7
2:00 271.3 F 29.1
3:00 272.3 F 29.9
4:00 271.5 F 29.2

Min 271.1 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 210.9 F 0.0
10:00 209.6 F 0.0
15:00 207.1 F 0.0
20:00 204.4 F 0.0
25:00 202.6 F 0.0
30:00 201.5 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 20:46
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

12/20/2021 08:00 AM
ACCEPTED: X
REJECTED:
Operator: U02