
Midmark M11 - v1.0.0
Total Cycles: 1380

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2021 04:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.0 F 0.0
2:00 178.1 F 3.0
4:00 185.0 F 6.2
6:00 211.2 F 10.2
8:00 223.2 F 5.2
10:00 230.9 F 8.9
12:00 239.8 F 11.5
14:00 239.7 F 11.4
16:00 238.6 F 9.9
18:00 251.3 F 16.3
20:00 262.5 F 22.8
22:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.5
2:00 272.1 F 29.1
3:00 271.2 F 28.6
4:00 271.2 F 29.0

Min 271.0 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 227.9 F 0.0
10:00 235.1 F 0.0
15:00 226.4 F 0.0
20:00 223.3 F 0.0
25:00 224.6 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

11/17/2021 05:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____