
Midmark M11 - v1.0.5
Total Cycles: 2396

Name: RITTER2
Sterilizer ID: V2313228
05/23/2023 07:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.3 F 0.0
2:00 85.9 F 0.0
4:00 133.7 F 0.0
6:00 110.9 F 0.0
8:00 146.8 F 1.4
10:00 176.8 F 4.5
12:00 211.4 F 8.2
14:00 206.3 F 12.5
16:00 218.4 F 5.3
18:00 229.0 F 9.3
20:00 238.6 F 12.1
22:00 239.4 F 11.0
24:00 237.0 F 9.1
26:00 250.1 F 15.5
28:00 261.2 F 21.8
30:00 271.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.3 F 28.8
2:00 272.7 F 29.3
3:00 271.9 F 28.8
4:00 272.1 F 28.7

Min 271.6 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.5 F 0.0
5:00 203.4 F 0.0
10:00 196.8 F 0.0
15:00 193.9 F 0.0
20:00 191.5 F 0.0
25:00 189.2 F 0.0
30:00 188.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:35
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:08:33
CYCLE COMPLETE

05/23/2023 08:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____