
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000923
13:04 10/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:21
PRESSURIZING 1:26
STERILIZING 2:31
136.0 C 229kPa 2:31
136.4 C 233kPa 3:01
136.3 C 233kPa 3:31
136.0 C 230kPa 4:01
136.3 C 233kPa 4:31
136.4 C 233kPa 5:01
136.3 C 233kPa 5:31
136.3 C 233kPa 6:01
136.3 C 233kPa 6:01
VENTING 6:01
AIR DRYING 6:54
CYCLE COMPLETE 15:47

Digital Signature #
01F056D3EB57A70E
