
Midmark M11 - v1.0.0
Total Cycles: 1780

Name: RITTER 1
Sterilizer ID: V2178198
04/01/2022 07:49 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	159.8 F	0.0
2:00	159.2 F	0.0
4:00	176.6 F	0.0
6:00	201.4 F	0.0
8:00	211.5 F	0.0
10:00	211.7 F	0.0
12:00	211.7 F	0.0
14:00	211.7 F	0.0
16:00	211.7 F	0.0
18:00	211.7 F	0.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 212F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

04/01/2022 08:09 AM