
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001363
13:57 24/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:48
STERILIZING 5:01
136.4 C 224kPa 5:01
136.1 C 223kPa 5:31
136.3 C 224kPa 6:01
136.3 C 224kPa 6:31
136.4 C 225kPa 7:01
136.2 C 221kPa 7:31
136.2 C 223kPa 8:01
136.2 C 224kPa 8:31
136.2 C 224kPa 8:31
VENTING 8:31
AIR DRYING 9:27
CYCLE COMPLETE 22:15

Digital Signature #
6DB5D962ADAFB508
